



— 314088 — 315031 — 314977

Date: 2017-02-16
 Scan Range: 400.0 nm to 700.0 nm
 Comments: 49002_Zeiss Axio_Un-Mounted
 ET - EGFP (FITC/Cy2)
 ET470/40x ET525/50m T495lpxr

Description
 ET470/40x
 ET525/50m
 T495LPXR

Batches / Filepath
 314088
 315031
 314977

ORIENTATION

Proper orientation of this filter is necessary in order to minimize autofluorescence and maximize performance. Please take a moment to locate the caret (arrow) located on the edge of each filter.

Excitation (x) filters should be positioned with the caret pointing toward the specimen and the inside of the cube, and away from the light source.

Emission (m) filters should be positioned with the caret pointing toward the specimen and the inside of the cube, and AWAY from the detector/eye.

Dichroic mirrors (bs) should be mounted with the coated surface TOWARD the light source, excitation filters and specimen. The coated side will be identified with a caret pointing to it, or a bevel. The beveled side is the smaller surface.

CLEANING AND HANDLING FILTERS

Handle coated pieces by the edges only. Clean gently only if necessary. Loose particles should be removed with a bulb puffer or filtered, pressurized air cleaner. If necessary, gently wipe surface using anhydrous alcohol and lint free lab towels. Use a new surface of towel with each wipe.

AVOID TOUCHING OR WIPING A/R COATED OR METAL MIRROR SURFACES

AVOID HANDLING EXPOSED COATINGS WITH BARE FINGERS



www.chroma.com